

**CERTIFICATE OF ANALYSIS****CheckLite 250 Plus**

Product code :61305

Lot No.: 20250912A

Expiry date: Jan. 2027

| <u>TEST PARAMETER</u>  | <u>RESULT</u> | <u>SPECIFICATION</u> |
|--|---------------|----------------------|
| Luciferin-luciferase reagent<br>Luminescence:Std.(%)   | 98.8          | 100±15               |
| Reconstitution buffer for luciferin-luciferase reagent<br>Background ATP( $\times 10^{-13}$ M) | 0.2           | $\leq 4.0$           |
| ATP releasing reagent<br>Background ATP( $\times 10^{-13}$ M)                                  | 0.1           | $\leq 4.0$           |

Storage condition : at 2-8°C, protected from light

This is to certify that above information is a true accurate copy of data obtained for this kit.

  
Hisako Kikuchi; ManagerDate Sep. 26, 2025KBG6031201ae-Eaa  
SA2509-0263Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company