

CERTIFICATE OF ANALYSIS**LuciPac A3 Surface**

Product code :60361

Lot No.: 20251002Y

Expiry date: Jan. 01, 2027

| <u>TEST PARAMETER</u> | <u>RESULT</u> | <u>SPECIFICATION</u> |
|--|---------------|-------------------------|
| When applied 100 μ L of 5pmol ADP Average Luminescence(RLU) | 12900 | $\geq 9100, \leq 16900$ |
| When applied 100 μ L of 5pmol AMP Average Luminescence(RLU) | 13200 | $\geq 9100, \leq 16900$ |

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Oct. 21, 2025KBG6036100ac-Eaa
SA2510-0199Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company