

CERTIFICATE OF ANALYSIS**Easy Plate AC-R**

Product code :61979

Lot No.: AR20250715A

Expiry date: Jul. 14, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	* See below
Spreading time	passed test	≦ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≦ 4 mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

*The ratio of the average colony count on the product to the plate count agar(%).

Escherichia coli (NBRC15034) : 50-200


Klebsiella pneumoniae (NBRC14940) : 50-200

Staphylococcus epidermidis (NBRC100911) : 50-200

Bacillus subtilis (NCTC10400) : 50-200

Latilactobacillus sakei (NBRC107868) : 50-200

Storage condition : 2-8°C


Hisako Kikuchi; Manager

Date Oct. 21, 2025

KBG6197900ab-Eaa
SA2510-0201

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company