

CERTIFICATE OF ANALYSIS**Easy Plate CC**

Product code :61974

Lot No.: CC20250801A

Expiry date: Jan. 31, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	≤ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤ 4 mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

Performance test

Escherichia coli (NBRC15034) : 70-200*

Klebsiella pneumoniae(NBRC14940) : 50-200*

Enterobacter cloacae (NBRC13536) : 50-200*

Citrobacter freundii (NBRC12681) : 50-200*

Staphylococcus aureus (ATCC25923) : No growth

Enterococcus faecalis (NBRC100482) : No growth

Proteus mirabilis (NBRC105697) : No growth

Staphylococcus xylosus (NBRC109770) : No growth

*The ratio of the average colony count on the product to the VRB agar medium(%).

Storage condition : 2-8°C


Hisako Kikuchi; ManagerDate Dec. 26, 2025