



CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20251223Y

Expiry date: Mar. 22, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	12200	\geq 9100, \leq 16900
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	12600	\geq 9100, \leq 16900

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Jan. 21, 2026

KBG6036100ac-Eaa
SA2601-0140

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company