



**CERTIFICATE OF ANALYSIS**

**LuciPac A3 Surface**

Product code :60361

Lot No.: 20251126Y

Expiry date: Feb. 25, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 $\mu$ L of 5pmol ADP Average Luminescence(RLU)	13300	$\geq 9100, \leq 16900$
When applied 100 $\mu$ L of 5pmol AMP Average Luminescence(RLU)	13600	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Hisako Kikuchi; Manager

Date Dec. 22, 2025

KBG6036100ac-Eaa  
SA2512-0348

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company