



CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20251127Y

Expiry date: Feb. 26, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100µL of 5pmol ADP Average Luminescence(RLU)	13300	≥9100, ≤16900
When applied 100µL of 5pmol AMP Average Luminescence(RLU)	13700	≥9100, ≤16900

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Hisako Kikuchi

Hisako Kikuchi; Manager

Date Dec. 22, 2025

KBG6036100ac-Eaa
SA2512-0349

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company