

CERTIFICATE OF ANALYSIS**LuciPac A3 Filter Assay**

Product code :60376

Lot No.: 20251114B

Expiry date: Feb. 10, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	13100	$\geq 9100, \leq 16900$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	13700	$\geq 9100, \leq 16900$

Measured with Lumitester Smart.

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Feb. 06, 2026KBG6037600ab-Eaa
SA2602-0067Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company