



CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20260210Y

Expiry date: May. 09, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	12500	\geq 9100, \leq 16900
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	12900	\geq 9100, \leq 16900

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Hisako Kikuchi
Hisako Kikuchi; Manager

Date Feb. 20, 2026

KBG6036100ac-Eaa
SA2602-0214

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company