



CERTIFICATE OF ANALYSIS

LuciPac A3 Surface (Pre-Moistened)

Product code :60367

Lot No.: 20260209M

Expiry date: Feb. 08, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence* (RLU)	17640	$\geq 10500, \leq 19500$

* When applied 100 μ L of AMP 2.5 pmol and ADP 2.5 pmol (total 5 pmol): measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Hisako Kikuchi
Hisako Kikuchi; Manager

Date Feb. 20, 2026

KBG6036700ab-Eaa
SA2602-0190

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company