



CERTIFICATE OF ANALYSIS

Histamine Check Swab

Product code :60448

Lot No.: 20251122J

Expiry date: May. 21, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Absorbance(600nm) of 100ppm histamine solution	1.94	$\geq 1.55, \leq 2.35$

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Hisako Kikuchi
Hisako Kikuchi; Manager

Date Feb. 20, 2026

KBG6044800ab-Eaa
SA2602-0189

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company