

CERTIFICATE OF ANALYSIS**Easy Plate SA**

Product code :61976

Lot No.: SA20260127A

Expiry date: Jan. 26, 2027

| <u>TEST PARAMETER</u> | <u>RESULT</u> | <u>SPECIFICATION</u> |
|----------------------------|---------------|----------------------|
| Performance test | passed test | See below |
| Spreading time | passed test | ≤ 60 sec |
| Gelling of medium | passed test | Gelling within 3 min |
| Curl of the base sheet | passed test | ≤ 4 mm |
| Appearance of culture area | passed test | Light yellow |
| Scanning of QR code | passed test | Passed |

Performance test

Staphylococcus aureus (ATCC25923) : 50-200 (as blue colony)^{*1}Staphylococcus aureus (NBRC13276) : 50-200 (as blue colony)^{*1}

Enterococcus faecalis (NBRC100482) : No growth

Bacillus subtilis (NBRC3134) : No growth

Escherichia coli (NBRC15034) : No growth

Proteus mirabilis (NBRC105697) : No growth

Bacillus licheniformis (NBRC12200) : Growth as pink colony

Staphylococcus epidermidis (NBRC100911) : ≤ 10 and no growth as blue colony^{*2}^{*1}The ratio of the average colony count on the product to the plate count agar (%).^{*2}The ratio of the average colony count on the product to Easy Plate AC (%).

Storage condition : 2-8°C



Hisako Kikuchi; Manager

Date Mar. 06, 2026KBG6197600aa-Eab
SA2603-0225Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company