

**CERTIFICATE OF ANALYSIS****Easy Plate AC-R**

Product code :61979

Lot No.: AR20251112A

Expiry date: Nov. 11, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	* See below
Spreading time	passed test	$\leq$ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	$\leq$ 4 mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

\*The ratio of the average colony count on the product to the plate count agar(%).

Escherichia coli (NBRC15034) : 50-200

Klebsiella pneumoniae (NBRC14940) : 50-200

Staphylococcus epidermidis (NBRC100911) : 50-200

Bacillus subtilis (NCTC10400) : 50-200

Latilactobacillus sakei (NBRC107868) : 50-200

Storage condition : 2-8°C

  
Hisako Kikuchi; Manager

Date Mar. 06, 2026

KBG6197900ab-Eaa  
SA2603-0228

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company