

CERTIFICATE OF ANALYSIS**LuciPac A3 Surface**

Product code :60361

Lot No.: 20260225Y

Expiry date: May. 24, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	12300	$\geq 9100, \leq 16900$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	13000	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Mar. 06, 2026KBG6036100ac-Eaa
SA2603-0219Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company