



**CERTIFICATE OF ANALYSIS**

**LuciPac A3 Water**

Product code :60365

Lot No.: 20251208W

Expiry date: Mar. 07, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 $\mu$ L of 5pmol ADP Average Luminescence(RLU)	13000	$\geq 9100, \leq 16900$
When applied 100 $\mu$ L of 5pmol AMP Average Luminescence(RLU)	13500	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Hisako Kikuchi; Manager

Date Mar. 06, 2026

KBG6036500ac-Eaa  
SA2603-0220

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company