



**CERTIFICATE OF ANALYSIS**

**LuciPac A3 Surface (Pre-Moistened)**

Product code :60367

Lot No.: 20260330M

Expiry date: Mar. 29, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence* (RLU)	17080	$\geq 10500, \leq 19500$

\* When applied 100 $\mu$ L of AMP 2.5 pmol and ADP 2.5 pmol (total 5 pmol): measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Hisako Kikuchi; Manager

Date Apr. 07, 2026

KBG6036700ab-Eaa  
SA2604-0075

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company