

CERTIFICATE OF ANALYSIS**Easy Plate EB**

Product code :61978

Lot No.: EB20260203A

Expiry date: Aug. 02, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	\leq 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	\leq 4 mm
Appearance of culture area	passed test	Orange to Yellow
Scanning of QR code	passed test	Passed

Performance test

Escherichia coli (NBRC15034) : 50-200*

Klebsiella pneumoniae(NBRC14940) : 50-200*

Proteus mirabilis (NBRC105697) : 50-200*

Salmonella enterica subsp. enterica (NBRC13245) : 50-200*

Pseudomonas aeruginosa (NBRC3899) : No growth

Enterococcus faecalis (NBRC100482) : No growth

*The ratio of the average colony count on the product to the plate count agar(%).

Storage condition : 2-8°C



Hisako Kikuchi; Manager

Date Apr. 07, 2026KBG6197800ab-Eaa
SA2604-0089Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company