

**CERTIFICATE OF ANALYSIS****Easy Plate LS**

Product code :61800

Lot No.: LS20260123A

Expiry date: Jul. 22, 2027

| <u>TEST PARAMETER</u>      | <u>RESULT</u> | <u>SPECIFICATION</u> |
|----------------------------|---------------|----------------------|
| Performance test           | passed test   | See below            |
| Spreading time             | passed test   | $\leq$ 60 sec        |
| Gelling of medium          | passed test   | Gelling within 3 min |
| Curl of the base sheet     | passed test   | $\leq$ 4 mm          |
| Appearance of culture area | passed test   | Light yellow         |
| Scanning of QR code        | passed test   | Passed               |

## Performance test

Listeria innocua (JCM 32814) : 33-200\*

Listeria monocytogenes serovar 4b (JCM 7675) : 33-200\*

Listeria monocytogenes serotype 1/2b (JCM 7676) : 33-200\*

Enterococcus faecalis (NBRC100482) : No growth

Escherichia coli (NBRC15034) : No growth

Bacillus licheniformis (NBRC12200) : No growth

\*The ratio of the average colony count on the product to the plate count agar(%).

Storage condition : 2-8°C

  
Hisako Kikuchi; ManagerDate Apr. 07, 2026KBG6180000aa-Eaa  
SA2604-0092Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company