



CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20260409Y

Expiry date: Jul. 08, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12300	≥9100, ≤16900
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12700	≥9100, ≤16900

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Apr. 20, 2026

KBG6036100ac-Eaa
SA2604-0255

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company