

## CERTIFICATE OF STERILITY AND QUALITY ASSURANCE

PRODUCT	LOT NUMBER	EXPIRY DATE	CERTIFICATION DATE
MW570PL 125	173103	31/03/2027	10/04/2026

### Listeria Isolation Transwab® - Plain Label

PARAMETER	SPECIFICATION	RESULT
Sterilisation Method	Irradiation	PASSED
Sterility Test	Visual inspection following incubation	PASSED
Performance Testing	Colour change induced by growth of following organisms: <i>Listeria monocytogenes</i> NCTC 5214 (24hr) <i>Listeria monocytogenes</i> NCTC 7993 (24hr) <i>Listeria monocytogenes</i> NCTC 11994 (24 hr) <i>Listeria innocua</i> NCTC 11288 (24 hr) No colour change induced by growth of following organisms <i>Staphylococcus aureus</i> NCTC 8532 (24hr) <i>Listeria grayii</i> NCTC 10815 (24 hr) <i>Enterococcus faecalis</i> ATCC® 19433 (24hr)	PASSED

Product has been manufactured in accordance with ISO 9001:2015

Performance testing conducted in accordance with: Internal Procedures

For and on behalf of Medical Wire & Equipment



Quality Assurance