



CERTIFICATE OF ANALYSIS

LuciSwab ES 2.8-3.2

Product code :60356

Lot No.: 25YE2212B03

| <u>TEST PARAMETER</u> | <u>RESULT</u> | <u>SPECIFICATION</u> |
|--------------------------------|---------------|----------------------|
| Blank Luminescence Assay (RLU) | | |
| Average Luminescence | passed test | ≤ 20 |
| Maximum Luminescence | passed test | ≤ 25 |

Measured with LuciPac A3 Water and Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Ryosuke Sugimoto; Manager

Date May. 25, 2026

KBG6035600ab-Eaa
SA2605-0343

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company