

**CERTIFICATE OF ANALYSIS****LuciSwab ES 3.7-4.2**

Product code :60357

Lot No.: 259E1Z18C01

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Blank Luminescence Assay (RLU)		
Average Luminescence	passed test	$\leq 20$
Maximum Luminescence	passed test	$\leq 25$

Measured with LuciPac A3 Water and Lumitester Smart

This is to that above information is a true and accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; ManagerDate May. 25, 2026KBG6035700ab-Eaa  
SA2605-0344Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company