

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20260611Y

Expiry date: Sep. 10, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12600	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	13100	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Hisako Kikuchi; Manager

Date Jun. 22, 2026

KBG6036100ac-Eaa
SA2606-0377

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company