

# CERTIFICATE OF ANALYSIS

## LuciPac A3 Water

Product code :60365

Lot No.: 20260612W

Expiry date: Sep. 11, 2027

| <u>TEST PARAMETER</u>                                        | <u>RESULT</u> | <u>SPECIFICATION</u>    |
|--------------------------------------------------------------|---------------|-------------------------|
| When applied 100μL of 5pmol ADP<br>Average Luminescence(RLU) | 13000         | $\geq 9100, \leq 16900$ |
| When applied 100μL of 5pmol AMP<br>Average Luminescence(RLU) | 13100         | $\geq 9100, \leq 16900$ |

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

*Hisako Kikuchi*  
Hisako Kikuchi; Manager

Date Jun. 22, 2026

KBG6036500ac-Eaa  
SA2606-0378

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company