

CERTIFICATE OF ANALYSIS

Easy Plate CC

Product code :61974

Lot No.: CC20260507A

Expiry date: Nov. 06, 2027


<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	≤ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤ 4 mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

Performance test

Escherichia coli (NBRC15034) : 70-200*
 Klebsiella pneumoniae(NBRC14940) : 50-200*
 Enterobacter cloacae (NBRC13536) : 50-200*
 Citrobacter freundii (NBRC12681) : 50-200*
 Staphylococcus aureus (ATCC25923) : No growth
 Enterococcus faecalis (NBRC100482) : No growth
 Proteus mirabilis (NBRC105697) : No growth
 Staphylococcus xylosus (NBRC109770) : No growth

*The ratio of the average colony count on the product to the VRB agar medium(%).

Storage condition : 2-8°C


 Hisako Kikuchi; Manager

Date Jun. 10, 2026

KBG6197400aa-Eab
 SA2606-0212

Quality Assurance Group, Edogawa
 Quality Unit
 Kikkoman Biochemifa Company