

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361


Lot No.: 20260630Y

Expiry date: Sep. 29, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	13100	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	14100	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Jul. 16, 2026

KBG6036100ac-Eaa
SA2607-0239

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company