

CERTIFICATE OF ANALYSIS**LuciPac A3 Surface (Pre-Moistened)**

Product code :60367

Lot No.: 20260615M

Expiry date: Jun. 14, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence* (RLU)	17880	≥ 10500 , ≤ 19500

* When applied 100 μ L of AMP 2.5 pmol and ADP 2.5 pmol (total 5 pmol): measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Hisako Kikuchi; Manager

Date Jul. 16, 2026KBG6036700ab-Eaa
SA2607-0242Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company