

CERTIFICATE OF ANALYSIS

LuciPac A3 Filter Assay

Product code :60376

Lot No.: 20260610A

Expiry date: Sep. 08, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	11600	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12200	$\geq 9100, \leq 16900$

Mesured with Lumitester Smart.

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Aug. 18, 2026

KBG6037600ab-Eaa
SA2608-0249

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company