

CERTIFICATE OF ANALYSIS

Easy Plate LS

Product code :61800

Lot No.: LS20260304A

Expiry date: Sep. 03, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	≤ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤ 4 mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

Performance test

Listeria innocua (JCM 32814) : 33-200*

Listeria monocytogenes serovar 4b (JCM 7675) : 33-200*

Listeria monocytogenes serotype 1/2b (JCM 7676) : 33-200*

Enterococcus faecalis (NBRC100482) : No growth

Escherichia coli (NBRC15034) : No growth

Bacillus licheniformis (NBRC12200) : No growth

*The ratio of the average colony count on the product to the plate count agar(%).

Storage condition : 2-8°C

J. Kikuchi
 Hisako Kikuchi; Manager

Date Aug. 18, 2026

KBG6180000aa-Eaa
SA2608-0255

Quality Assurance Group, Edogawa
 Quality Unit
 Kikkoman Biochemifa Company