

## CERTIFICATE OF ANALYSIS

### Easy Plate EC

Product code :61975

Lot No.: EC20260810A

Expiry date: Feb. 09, 2028

| <u>TEST PARAMETER</u>      | <u>RESULT</u> | <u>SPECIFICATION</u> |
|----------------------------|---------------|----------------------|
| Performance test           | passed test   | See below            |
| Spreading time             | passed test   | $\leq 60$ sec        |
| Gelling of medium          | passed test   | Gelling within 3 min |
| Curl of the base sheet     | passed test   | $\leq 4$ mm          |
| Appearance of culture area | passed test   | Light yellow         |
| Scanning of QR code        | passed test   | Passed               |

#### Performance test

Escherichia coli (NBRC15034) : 70-200\*

Klebsiella pneumoniae(NBRC14940) : 50-200\*

Enterobacter cloacae (NBRC13536) : 50-200\*

Citrobacter freundii (NBRC12681) : 50-200\*

Staphylococcus aureus (ATCC25923) : No growth

Enterococcus faecalis (NBRC100482) : No growth

Proteus mirabilis (NBRC105697) : No growth

Staphylococcus xylosus (NBRC109770) : No growth

\*The ratio of the average colony count on the product to the VRB agar medium(%).

Storage condition : 2-8°C



Hisako Kikuchi; Manager

Date Sep. 29, 2026

KBG6197500aa-Eab  
SA2609-0389

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company