

CERTIFICATE OF ANALYSIS**LuciSwab ES 2.0-2.2**

Product code :60355

Lot No.: 25YE3212A01

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Blank Luminescence Assay (RLU)		
Average Luminescence	passed test	≤ 20
Maximum Luminescence	passed test	≤ 25

Measured with LuciPac A3 Water and Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Oct. 13, 2026KBG6035500ab-Eaa
SA2610-0184Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company