kikkoman Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface (Pre-Moistened)

Product code :60367

Lot No.: 20260914M Expiry date: Sep. 13, 2027

TEST PARAMETER	RESULT	SPECIFICATION
Average Luminescence*(RLU)	17190	≧10500, ≦19500

* When applied 100μL of AMP 2.5 pmol and ADP 2.5 pmol (total 5 pmol): measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H. Kikuchi

Hisako Kikuchi; Manager

Date Oct. 13, 2026

KBG6036700ab-Eaa
SA2610-0175

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company