

## CERTIFICATE OF ANALYSIS

### Histamine Test

Product code :61341

Lot No.: 20260818A

Expiry date: Mar. 2029

| <u>TEST PARAMETER</u>                                                               | <u>RESULT</u> | <u>SPECIFICATION</u> |
|-------------------------------------------------------------------------------------|---------------|----------------------|
| Quantifiability of histamine(%)<br>(Ratio to liquid preparation)                    | 100.6         | 100±5                |
| Absorbance of reagent blank                                                         | 0.011         | ≤0.04                |
| Histamine dehydrogenase<br>activity(U/vial)                                         | 2.19          | ≥ 0.64               |
| Concentration of Histamine std.<br>solution(%)<br>(Ratio to the reference solution) | 100.1         | 100±5                |

Storage condition : at 2-8°C protected from light

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Hisako Kikuchi; Manager

Date Oct. 13, 2026

KBG6044101aa-Eaa  
SA2610-0177

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company