



Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

LuciPac A3 Water

Product code :60365

Lot No.: 20260908W

Expiry date: Dec. 07, 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	11800	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	11900	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Sep. 14, 2026

KBG6036500ac-Eaa
SA2609-0240

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company