kikkoman Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20261014Y Expiry date: Jan. 13, 2028

TEST PARAMETER	RESULT	SPECIFICATION
When applied 100µL of 5pmol ADP Average Luminescence(RLU)	11800	≧9100, ≦16900
When applied 100µL of 5pmol AMP Average Luminescence(RLU)	12200	≧9100, ≦16900

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H. Kikuchi

Hisako Kikuchi; Manager

Date Oct. 26, 2026

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company

KBG6036100ac-Eaa
SA2610-0408