

CERTIFICATE OF ANALYSIS

Easy Plate EB

Product code :61978

Lot No.: EB20260731B

Expiry date: Jan. 30, 2028

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	≤ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤ 4 mm
Appearance of culture area	passed test	Orange to Yellow
Scanning of QR code	passed test	Passed

Performance test

Escherichia coli (NBRC15034) : 50-200^{*}

Klebsiella pneumoniae(NBRC14940) : 50-200^{*}

Proteus mirabilis (NBRC105697) : 50-200^{*}


Salmonella enterica subsp. enterica (NBRC13245) : 50-200^{*}

Pseudomonas aeruginosa (NBRC3899) : No growth

Enterococcus faecalis (NBRC100482) : No growth

^{*}The ratio of the average colony count on the product to the plate count agar(%).

Storage condition : 2-8°C


 Hisako Kikuchi; Manager

Date Oct. 26, 2026

KBG6197800ab-Eaa
SA2610-0414

Quality Assurance Group, Edogawa
 Quality Unit
 Kikkoman Biochemifa Company