kikkoman Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

Easy Plate EC

Product code :61975

Lot No.: EC20260910A Expiry date: Mar. 09, 2028

TEST PARAMETER	RESULT	SPECIFICATION
Performance test	passed test	See below
Spreading time	passed test	≦ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≦ 4 mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

Performance test

Escherichia coli (NBRC15034) : 70-200*
Klebsiella pneumoniae(NBRC14940) : 50-200*
Enterobacter cloacae (NBRC13536) : 50-200*
Citrobacter freundii (NBRC12681) : 50-200*
Staphylococcus aureus (ATCC25923) : No growth
Enterococcus faecalis (NBRC100482) : No growth
Proteus mirabilis (NBRC105697) : No growth
Staphylococcus xylosus (NBRC109770) : No growth

*The ratio of the average colony count on the product to the VRB agar medium(%).

Storage condition : 2-8℃

H. Kikuchi

Hisako Kikuchi; Manager

Date Oct. 26, 2026

KBG6197500aa-Eab
SA2610-0415

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company