



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL: +81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

June 24, 2024

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Beta-lactoglobulin ELISA Kit II	24JUSFBL056	JUN.19 2025

The product kit, batch 24JUSFBL056, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : June 24, 2024)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.009	0.008	0.008	0.009	
0.78	0.043	0.040	0.040	0.040	0.041	3.7
1.56	0.080	0.077	0.075	0.073	0.076	3.9
3.13	0.148	0.143	0.141	0.140	0.143	2.5
6.25	0.289	0.280	0.279	0.264	0.278	3.7
12.5	0.559	0.544	0.543	0.524	0.543	2.6
25	1.048	1.030	1.022	1.015	1.029	1.4
50	1.896	1.871	1.859	1.839	1.866	1.3

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.7	7.0-13.0
B	6.2	4.3-8.0
C	3.2	2.6-4.8