



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

February 22, 2024

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Beta-lactoglobulin ELISA Kit II	24FESFBL055	FEB.19 2025

The product kit, batch 24FESFBL055, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : February 22, 2024)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.007	0.007	0.007	0.008	0.007	
0.78	0.038	0.041	0.040	0.041	0.040	3.5
1.56	0.078	0.076	0.076	0.079	0.077	1.9
3.13	0.144	0.140	0.138	0.140	0.141	1.8
6.25	0.267	0.263	0.266	0.263	0.265	0.8
12.5	0.523	0.515	0.510	0.514	0.516	1.1
25	0.982	0.968	0.945	0.977	0.968	1.7
50	1.786	1.784	1.763	1.795	1.782	0.8

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.0	7.0-13.0
B	5.9	4.3-8.0
C	3.3	2.6-4.8