



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL: +81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

August 5, 2022

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

### Product:

Product name	Batch number	Expiration date
Beta-lactoglobulin ELISA Kit II	22AUSFBL048	AUG.04 2023

The product kit, batch 22AUSFBL048, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : August 5, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.012	0.011	0.010	0.012	0.011	
0.78	0.044	0.043	0.043	0.042	0.043	1.9
1.56	0.078	0.076	0.076	0.074	0.076	2.1
3.13	0.139	0.138	0.138	0.135	0.138	1.3
6.25	0.271	0.265	0.269	0.261	0.267	1.7
12.5	0.527	0.525	0.523	0.511	0.522	1.4
25	1.012	1.015	1.018	0.991	1.009	1.2
50	1.885	1.929	1.940	1.896	1.913	1.4

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	9.3	7.0-13.0
B	4.4	4.3-8.0
C	3.0	2.6-4.8