



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL: +81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

May 8, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Beta-lactoglobulin ELISA Kit II	20MYSFBL036	MAY. 07 2021

The product kit, batch 20MYSFBL036, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : May 8, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.012	0.013	0.012	0.013	0.013	
0.78	0.050	0.049	0.049	0.050	0.050	1.2
1.56	0.085	0.086	0.084	0.084	0.085	1.1
3.12	0.156	0.153	0.150	0.154	0.153	1.6
6.25	0.297	0.299	0.289	0.279	0.291	3.1
12.5	0.566	0.555	0.544	0.555	0.555	1.6
25	1.076	1.060	1.042	1.057	1.059	1.3
50	1.977	1.968	1.883	1.937	1.941	2.2

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	6.7	5.7-10.6
B	4.3	3.7-6.8
C	2.6	2.2-4.2