



## Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,  
Yokohama-shi, 230-8504, Japan  
TEL:+81-45-586-2514  
E-mail: info\_miobs\_e@morinaga.co.jp

May 21, 2024

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

### Product:

Product name	Batch number	Expiration date
Total Milk ELISA Kit II	24MYSFTM013	MAY.19 2025

The product kit, batch 24MYSFTM013, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date: May 21, 2024)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.010	0.010	0.010	0.010	
0.78	0.052	0.050	0.051	0.051	0.051	1.6
1.56	0.096	0.097	0.099	0.097	0.097	1.3
3.13	0.174	0.174	0.171	0.171	0.173	1.0
6.25	0.331	0.327	0.326	0.322	0.327	1.1
12.5	0.625	0.619	0.620	0.612	0.619	0.9
25	1.147	1.161	1.141	1.143	1.148	0.8
50	1.961	1.939	1.927	1.922	1.937	0.9

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.8	7.4-13.7
B	6.3	4.9-9.0
C	3.9	3.7-6.9