



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL: +81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

December 11, 2023

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Total Milk ELISA Kit II	23DESFTM009	DEC.07 2024

The product kit, batch 23DESFTM009, passed the quality control.

(1) Absorbance of the standard curve

(Assay date: December 11, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.011	0.009	0.009	0.010	0.010	
0.78	0.051	0.051	0.049	0.049	0.050	2.3
1.56	0.094	0.092	0.090	0.089	0.091	2.4
3.13	0.179	0.171	0.172	0.172	0.174	2.1
6.25	0.343	0.329	0.328	0.333	0.333	2.1
12.5	0.654	0.639	0.637	0.650	0.645	1.3
25	1.233	1.207	1.198	1.199	1.209	1.4
50	2.162	2.117	2.121	2.119	2.130	1.0

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.6	7.4-13.7
B	7.2	4.9-9.0
C	4.3	3.7-6.9