



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL: +81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

August 10, 2023

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Total Milk ELISA Kit II	23AUSFTM006	AUG.07 2024

The product kit, batch 23AUSFTM006, passed the quality control.

(1) Absorbance of the standard curve

(Assay date: August 10, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.012	0.012	0.013	0.012	0.012	
0.78	0.056	0.056	0.051	0.050	0.053	6.0
1.56	0.095	0.095	0.094	0.093	0.094	1.0
3.13	0.179	0.174	0.177	0.175	0.176	1.3
6.25	0.336	0.337	0.332	0.328	0.333	1.2
12.5	0.647	0.631	0.627	0.638	0.636	1.4
25	1.181	1.184	1.186	1.187	1.185	0.2
50	1.970	2.080	2.066	2.052	2.042	2.4

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.4	7.4-13.7
B	6.4	4.9-9.0
C	3.8	3.7-6.9