



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

February 24, 2023

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Total Milk ELISA Kit II	23FESFTM001	FEB.22 2024

The product kit, batch 23FESFTM001, passed the quality control.

(1) Absorbance of the standard curve

(Assay date: February 24, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.011	0.009	0.010	0.011	0.010	
0.78	0.059	0.057	0.057	0.056	0.057	2.2
1.56	0.103	0.100	0.099	0.098	0.100	2.2
3.13	0.199	0.191	0.187	0.186	0.191	3.1
6.25	0.373	0.363	0.362	0.359	0.364	1.7
12.5	0.713	0.689	0.685	0.687	0.694	1.9
25	1.323	1.273	1.272	1.266	1.284	2.1
50	2.262	2.230	2.215	2.215	2.231	1.0

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.8	7.4-13.7
B	7.3	4.9-9.0
C	5.3	3.7-6.9