



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

April 19, 2024

Quality Control Report

Morinaga Institute of Biological Science, Inc._

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	24APSFSY069	APR.17 2025

The product kit, batch 24APSFSY069, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : April 19, 2024)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.012	0.011	0.007	0.009	0.010	
0.78	0.036	0.038	0.037	0.037	0.037	2.2
1.56	0.072	0.076	0.075	0.077	0.075	2.9
3.13	0.145	0.143	0.141	0.140	0.142	1.6
6.25	0.274	0.271	0.265	0.269	0.270	1.4
12.5	0.516	0.507	0.516	0.514	0.513	0.8
25	0.965	0.957	0.956	0.973	0.963	0.8
50	1.715	1.773	1.734	1.808	1.758	2.4

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.1	6.6-12.3
B	13.4	9.4-17.5
C	6.5	4.8-8.9