



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

January 26, 2024

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	24JASFSY068	JAN.24 2025

The product kit, batch 24JASFSY068, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : January 26, 2024)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.006	0.006	0.006	0.007	0.006	
0.78	0.047	0.046	0.045	0.045	0.046	2.1
1.56	0.086	0.079	0.083	0.083	0.083	3.5
3.13	0.155	0.147	0.151	0.149	0.151	2.3
6.25	0.306	0.275	0.287	0.267	0.284	6.0
12.5	0.593	0.557	0.568	0.553	0.568	3.2
25	1.105	1.068	1.069	1.028	1.068	2.9
50	1.997	1.978	1.971	1.960	1.977	0.8

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.2	6.6-12.3
B	13.8	9.4-17.5
C	6.3	4.8-8.9