



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL: +81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

September 7, 2023

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	23SESFSY066	SEP.05 2024

The product kit, batch 23SESFSY066, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : September 7, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.007	0.006	0.006	0.006	0.006	
0.78	0.044	0.043	0.044	0.041	0.043	3.3
1.56	0.085	0.083	0.082	0.080	0.083	2.5
3.13	0.161	0.155	0.158	0.151	0.156	2.7
6.25	0.310	0.299	0.303	0.287	0.300	3.2
12.5	0.576	0.574	0.568	0.552	0.568	1.9
25	1.076	1.043	1.045	1.034	1.050	1.7
50	1.938	1.924	1.924	1.886	1.918	1.2

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.8	6.2-11.6
B	12.2	8.5-15.7
C	6.8	4.4-8.2