



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

May 19, 2023

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	23MYSFSY064	MAY.18 2024

The product kit, batch 23MYSFSY064, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : May 19, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.007	0.006	0.007	0.007	0.007	
0.78	0.048	0.045	0.044	0.042	0.045	5.6
1.56	0.090	0.083	0.081	0.079	0.083	5.8
3.13	0.163	0.155	0.157	0.145	0.155	4.8
6.25	0.311	0.298	0.302	0.289	0.300	3.0
12.5	0.588	0.566	0.569	0.557	0.570	2.3
25	1.046	1.051	1.033	1.031	1.040	0.9
50	1.835	1.931	1.931	1.895	1.898	2.4

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.2	6.2-10.1
B	12.9	8.5-13.8
C	6.3	4.4-7.8