



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

October 11, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	22OCSFSY061	OCT.07 2023

The product kit, batch 22OCSFSY061, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : October 11, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.008	0.006	0.007	0.008	0.007	
0.78	0.053	0.052	0.051	0.052	0.052	1.6
1.56	0.097	0.095	0.095	0.094	0.095	1.3
3.13	0.178	0.173	0.175	0.176	0.176	1.2
6.25	0.345	0.339	0.339	0.334	0.339	1.3
12.5	0.651	0.647	0.655	0.646	0.650	0.6
25	1.229	1.223	1.202	1.202	1.214	1.2
50	2.123	2.141	2.163	2.146	2.143	0.8

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.3	6.2-10.1
B	10.7	8.5-13.8
C	5.9	4.4-7.8